



NATIONAL TYPE EVALUATION PROGRAM

Certificate of Conformance

for Weighing and Measuring Devices

For:
Non-Computing Scale
Digital Electronic
Model: AGS Series
 n_{max} : 6000
Accuracy Class: III

Submitted By:
Universal Weight Electronic Co., Ltd.
4th Floor, No. 53
Baoxing Road / Xindian District
New Taipei City 231
Taiwan (R.O.C.)
Tel: +866-2-29180121
Contact: Jeffrey Lu
Email: Jeffrey@uwe.com.tw

Standard Features and Options

- Automatic Zero Tracking (AZT)
- Initial Zero Setting Mechanism (IZSM)
- Semi-Automatic Zero (Push Button)
- Semi-Automatic Tare (Push Button)
- Multi-interval
- AC Power (120V)
- DC Power (4.5V, 6V)
- Liquid Crystal Display
- RS 232 Communication Port
- USB Communication Port

Model	Capacity	d	n_{max}
AGS-300	150 g / 300 g	0.05 g / 0.1 g	3000 / 3000
AGS-600	300 g / 600 g	0.1 g / 0.2 g	3000 / 3000
AGS-1500	600 g / 1500 g	0.2 g / 0.5 g	3000 / 3000
AGS-3000	3000 g 6.6 lb	0.5 g 0.002 lb	6000 3300
AGS-6000	6000 g 13.2 lb	1 g 0.005 lb	6000 2640
AGS-12K	12 kg 26.4 lb	0.002 kg 0.005 lb	6000 5280
AGS-30K	30 kg 66 lb	0.005 kg 0.02 lb	6000 3300

Load Cells Used: For AGS -300 to -600 Minbea model BCL (Non-NTEP) and for AGS-6000 to -30K True-Tec PA06 (NTEP CC 16-098)

Temperature Range: 0 °C to 40 °C (32 °F to 104 °F)

This device was evaluated under the National Type Evaluation Program and was found to comply with the applicable technical requirements of *Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices*. Evaluation results and device characteristics necessary for inspection and use in commerce are on the following pages. *Editorial changes, not affecting the type or metrological content, corrected this certificate.

Hal Prince
Chairman, NCWM, Inc.

Craig VanBuren
Chair, NTEP Committee
Issued: October 30, 2020

1135 M Street, Suite 110 / Lincoln, Nebraska 68508

The National Conference on Weights and Measures (NCWM) does not approve, recommend or endorse any proprietary product or material, either as a single item or as a class or group. Results shall not be used in advertising or sales promotion to indicate explicit or implicit endorsement of the product or material by the NCWM.



Universal Weight Electronic Co., Ltd.
Non-Computing Scale / AGS Series

Application: Non-computing scale for general purpose weighing applications.

Identification: The required identification markings are located under the platter

Sealing: A wire security seal is threaded through a post on the bottom of the scale that secure a plate to prevent access to the jumpers.

Test Conditions: This Certificate of Conformance supersedes Certificate of Conformance Number 18-130 and was issued to indicate a transfer of ownership from Intelligent Weighing Technology to Universal Weight Electronics Co., Ltd. No additional testing was required. Test Conditions are listed below for reference.

Certificate of Conformance Number 18-130: The emphasis of the evaluation was on the device design, marking, operation, performance, and compliance with influence factor requirements. Three scales were provided for this evaluation: an AGS-300 with a 150 g / 300 g x 0.5 g / 0.1 g; an AGS-3000 with a 3000 g (6.6 lb) x 0.5 g (0.002 lb) capacity; and an AGS-30K with a 30 kg (66 lb) x 0.005 kg (0.02 lb) capacity. A series of increasing/decreasing load tests, discrimination tests and eccentricity tests were performed. The scales were tested with an AC power supply of 102 VAC and 132 VAC and also with a DC power supply of 2.5 VDC and 5.0 VDC on the AGS-300 and 4.7 VDC and 6.6VDC on the AGS-3000 and AGS-30K . A warm up test, power interruption test, and electromagnetism tests were also performed on each. The scales were tested over a temperature range of 0°C to 40°C (32° F to 104°F). A load of approximately one-half capacity was applied to each scale over 100 000 times. The scales were tested periodically during this time.

Evaluated By: E. Morabito (NY) 18-130; M. Manheim (NCWM) 18-130A1

Type Evaluation Criteria Used: *Handbook 44 Specifications, Tolerances, and Other Technical Requirements for Weighing and Measuring Devices*, 2018 Edition. *NCWM Publication 14: Measuring Devices*, 2018 Edition.

Conclusion: The results of the evaluation and information provided by the manufacturer indicate the device complies with applicable requirements.

Information Reviewed By: J. Truex (NCWM) 18-130; D. Flocken (NCWM) 18-130A1

Example(s) of Device:



AGS-300

AGS-3000

AGS-30K

Method of Sealing